



DECLARATION OF CONFORMITY

Product **Laboratory centrifuge**

Models **Harrier, Harrier R, Hawk, Hawk R, MicroCentaur R**

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product classification on the basis of the Directive 2014/35/EU(LVD)

Product complies with the requirements, including the requirements of harmonised standards:

EN 980:2008

EN ISO 18113-3:2011

EN 13612:2006

EN 61326-2-6:2006

EN 13612:2002/AC:2002 EN

EN 61010-2-101:2002

ISO 14971:2012

EN 62304:2006

EN ISO 18113-1:2011

EN 62304:2006/AC:2008

EN 61010-1:2010

EN 62366:2008

EN 61010-2-020:2006

The products comply with Directive 2014/30/EU (EMC) and Directive 2011/65/EU (RoHS2)

Heathfield
5th September 2018

Date of Issue

MSE Centrifuges Limited
11320509

11 Browning Road,
Heathfield TN21 8DB

*Quality policy in line with ISO
9001:2015 Certifying authority*

Patricia Sturgess
Managing Director

Steve Marriott
Technical Manager

